


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-IP*

S.N.: *529*

Date: *MAY 30 81*



a) Output signal amplitude:

> 30 VOLTS

b) Pulse width:

≤ 1 NSEC

c) Jitter, stability:

OK

d) Prime power:

+15 ± 1.0 V

165 mA (max)

[Signature]

1.0 NSEC/DIV
32 VOLTS/DIV