


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVG-1*

S.N.: *530*

Date: *MAY 30 81*



a) Output signal amplitude:

$\geq +250 \text{ V}$

b) Pulse width:

$\leq 4 \text{ NS}$

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0 \text{ V}$

250 mA (max)

5 NS/DIV

100 VOLTS/DIV

40 KHZ.