

PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVO-P


S.N.: 539

Date: JUNE 30 81

- a) Output signal amplitude:  
10 TO 50 AMPS
- b) Pulse width:  
25 NSEC
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 1 KHZ.
- f) Jitter, stability:  
OK
- g) Prime power:

+15 ± 1.0 V

175 mA (MAX)



20 AMPS/DIV  
10 NSEC/DIV  
1 KHZ