


PULSE GENERATOR

PERFORMANCE CHECK

Model: **AUP-AV-1-N-TLL-05**

S.N.: **545**

Date: **JUNE 18 81**



a) Output signal amplitude:

0 TO - 10V

b) Pulse width:

0 TO 4 nSEC

c) Rise time:

≤ 100 nSEC

d) Fall time:

≤ 135 nSEC

e) PRF:

0 TO 1.0 MHz

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 V

225 mA (MAX)

10 V / DIV

1.0 nSEC / DIV

400 kHz