


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-S-1-IP*

S.N.: *546*

Date: *JUNE 11 81*



*1.0 NSEC/DIV  
10 VOLTS/DIV  
400 KHZ*

a) Output signal amplitude:

*> 10 V*

b) Pulse width:

*≤ 150 PSEC*


c) Jitter, stability:

*OK*

d) Prime power:

*+15 ± 1.0 VOLT*

*200 mA (MAX)*



*100 PSEC/DIV  
10 VOLTS/DIV  
400 KHZ*

*[Handwritten signature]*