


IMPULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-5

S.N.: 547

Date: June 11 81



0.5 nsec/div
10 volts/div

a) Output signal amplitude:

+15 V

b) Pulse width:

≤ 500 psec

c) Jitter, stability:

OK

d) Prime power:

+15 ± 1.0 VOLT

170 mA (MAX)

[Signature]