


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-5-1-05-1V*

S.N.: *549*

Date: *6427 27 81*



*1.0 μ SEC/DIV
10 VOLTS/DIV
400 KHZ*

a) Output signal amplitude:

-10V

b) Pulse width:

≤ 150 psec

c) Jitter, stability:


OK.

d) Prime power:

+15 \pm 1.0V

200 mA (MAX)

AK



*100 PSEC/DIV
10 VOLTS/DIV
400 KHZ.*