

PULSE GENERATOR
PERFORMANCE CHECK

Model: **PAVI***
S.N.: **553**
Date: **JUN 27 81**

Mod v ad

- a) Output signal amplitude:
0 TO +40V
- b) Pulse width:
1 TO 100 NSEC.
- c) Rise time: **SEE FIG 1, +2**
≤ 300 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
SEE FIG 2.
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 V
165 mA.

32 VOLTS/DIV
5 NSEC/DIV
20 kHz.

[Signature]

* WITH ELECTRONIC
AMPLITUDE CONTROL
OPTION,