

IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-MPT-2*

S.N.: *566*

Date: *JULY 27 81*

a) Output signal amplitude:

> 60 V

b) Pulse width:

*> 2.0 nsec **


c) Jitter, stability:

OK

d) Prime power:

15 = 1.0 VOLTS

180 mA.



32 VOLTS/DIV

2.0 nsec/DIV

400 kHz

** SPECIAL UNIT.*

PW = 2 nsec AT

FWHM RATHER

*TITAN AT 10%
RISE POINT.*