

IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-MPT-2*

S.N.: *567*

Date: *JULY 27 81.*

a) Output signal amplitude:

≥ +60 VOLTS

b) Pulse width:

≥ 2.0 NSEC

c) Jitter, stability:

OK

d) Prime power:

+15 ± 1.0 VOLTS

190 mA (MAX)


** SPECIFIC UNIT.*

PW = 2 NSEC

AT FWHM

RATHER THAN AT

10% RISE POINT



*32 VOLTS/DIV
2.0 NSEC/DIV
400 KHZ*