


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVP-AU-1-P*

S.N.: *577*

Date: *April 31 81*



*1.0 nsec/div
10 volts/div
400 kHz*



*100 psec/div
2 volts/div
400 kHz*

- a) Output signal amplitude:
0 TO + 10 V
- b) Pulse width:
0 TO 4 nsec
- c) Rise time:
≤ 100 psec
- d) Fall time:
≤ 135 psec
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK.
- g) Prime power:
*+15 ± 1.0 V
340 mA MAX*

[Signature]