


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-MU-2-05-N*

S.N.: *582*

Date: *AUG 18 81*



a) Output signal amplitude:

*21 - 40V*

b) Pulse width:

*600 PSEC (10% RISE)*

*390 PSEC (FWHM)*

c) Jitter, stability:

*OK*

d) Prime power:

*+15 ± 1.0 V*

*250 mA (max)*

*40 VOLT/DIV*

*1.0 NSEC/DIV*

*40 KHZ.*

e) PRF:

*0 TO 500 KHZ*