


PULSE GENERATOR

PERFORMANCE CHECK

Model: **AVM-2-N-M1**

S.N.: **583**

Date: **SEPT 181**



a) Output signal amplitude:

**0 TO -15 V**

b) Pulse width\*

**0 TO 10 NSEC**

c) Rise time:

**≤ 100 PSEC.**

d) Fall time\*

**≤ 400 PSEC**

e) PRF:\*

**0 TO 1 MHz**

f) Jitter, stability:

**OK**

g) Prime power:

**+24 ± 1.0 VOLT**  
**400 mA (MAX)**

**1 NSEC/DIV**  
**10 VOLTS/DIV**  
**1 MHz**

**\* SPECIAL UNIT,  
NON-STANDARD  
VALUES.**