


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVK-A-P-05*

S.N.: *584*

Date: *SEPT 181*



*1 NSEC/DIV  
10 VOLTS/DIV  
20 KHZ.*

- a) Output signal amplitude:  
*0 TO +15V*
- b) Pulse width:  
*4 TO 100 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*SEE FIG 2 + FIG 1*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+15 ± 1.0 VOLT.*

*225 mA (MAX)*