


PULSE GENERATOR

PERFORMANCE CHECK


Model: **AVK-V-N-TTL-05**

S.N.: **586**

Date: **AUG 18 81**



**10 VOLTS/DIV
10 NSEC/DIV
40 KHZ**



**10 VOLTS/DIV
1.0 NSEC/DIV
40 KHZ**

a) Output signal amplitude:

$\geq -5V$

b) Pulse width:

0 TO 50 NSEC.

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 2 NSEC

e) PRF:

0 TO 50 KHZ

f) Jitter, stability:

OK

g) Prime power:

$+15 \pm 1.0$ VOLT.

225 mA (MAX.)