


PULSE GENERATOR
PERFORMANCE CHECK

Model: **AVP-AU-1-P**
S.N.: **589**
Date: **SEPT 30 81**



- a) Output signal amplitude:
0 TO +10 V
-) Pulse width:
0 TO 4.0 NS
-) Rise time:
≤ 100 PSEC
-) Fall time:
≤ 135 PSEC
-) PRF:
≤ 1.0 MHz
-) Jitter, stability:
OK
-) Prime power:

+15 ± 1.0 V

1.0 NS/CM
10 VOLTS/CM
400 KHZ

235 mA (MAX)