


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-M-1-ECL-P*

S.N.: *599*

Date: *SEPT 30 81*



a) Output signal amplitude:

0 TO +10V

b) Pulse width:

0 TO 4.0 NS

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 135 PSEC

e) PRF:

0 TO 1.0 MHz

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 VOLTS

235 mA (MAX)

1.0 NS/DIV
10 VOLTS/DIV
400 KHz.