


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-2-P*

S.N.: *602*

Date: *SEPT 28 81*



*10 AMPS/DIV
5 NSEC/DIV
20 KHZ.*

- a) Output signal amplitude:
0 TO +10 AMPS
- b) Pulse width:
0 TO 15 NSEC
- c) Rise time:
≤ 1 NSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK.
- g) Prime power:

+15 ± 1.0 V

300 mA MAX

Affo