


PULSE GENERATOR
PERFORMANCE CHECK

Model: **AVM-2-P-D-PS**

S.N.: **608**

Date: **SEPT 14 81**



1.0 NS/DIV
10 VOLTS/DIV
4.0 MHz

- a) Output signal amplitude:
0 TO +15 V
- b) Pulse width:
0 TO 2 NS
- c) Rise time:
≤ 100 PSE
- d) Fall time:
≤ 135 PSE
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:*

110/220 V 50-60 Hz

* THIS UNIT SHIPPS
CONNECTED FOR
270 V OPERATION.

Handwritten signature