


PULSE GENERATOR
PERFORMANCE CHECK

Model: **AVK-P-TTK**
S.N.: **621**
Date: **OCT 26 81**



- a) Output signal amplitude:
 $\geq +5$ VOLTS
- b) Pulse width:
4 TO 100 NS
- c) Rise time:
 ≤ 100 PSEC
- d) Fall time:
 ≤ 2 NS
- e) PRF:
SEE FIG 1 & 2
- f) Jitter, stability:
OK
- g) Prime power:
 $+15 \pm 1.0$ VOLTS
250 mA (MAX)

5 VOLT / DIV
1.0 NS / DIV
40 KITZ

[Handwritten Signature]