


PULSE GENERATOR
PERFORMANCE CHECK

Model: **AVM-3-IP**
S.N.: **623**
Date: **NOV 23 61**



1.0 NSEC/DIV
10 VOLTS/DIV
400 KHZ

-) Output signal amplitude:
0 to +15 V
-) Pulse width:
0 to 15 NSEC
-) Rise time:
≤ 150 PSEC
-) Fall time:
≤ 600 PSEC
-) PRF:
0 to 25 MHz
-) Jitter, stability:
OK
-) Prime power:
+24 to +30 V DC
400 mA.

[Handwritten signature]