


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVL-AV-P*  
S.N.: *631*  
Date: *OCT 26 81*



- a) Output signal amplitude:  
*0 TO + 250 V*
- b) Pulse width:  
*0 TO 100 nsec*
- c) Rise time:  
*≤ 5 nsec*
- d) Fall time:  
*≤ 5 nsec*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 ± 1.0 VOLTS*  
*400 mA (MAX)*

*10 nsec/DIV*  
*≈ 160 VOLTS/DIV*  
*2 KHz*