


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-5-1-A-P*

S.N.: *632*

Date:



Output signal amplitude:

*≥ 10 VOLT*

Pulse width:

*≤ 135 PSEC*

Jitter, stability:

*OK.*

Prime power:

*± 15 ± 1.0 VOLT*

*180 mA*

*10 VOLT/DIV*

*1.0 NSEC/DIV*

*400 KHZ*