


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *MNH-5-1-05-N*

S.N.: *637*

Date: *NOV 13 81*



*100 PSEC/DIV*  
*10 VOLTS/DIV*  
*400 KHZ*

a) Output signal amplitude:

*± 10 V*

b) Pulse width:

*± 135 PSEC*


c) Jitter, stability:

*OK*

d) Prime power:

*+15 ± 1.0 VOLT*

*175 MA*



*1.0 NSEC/DIV*  
*10 VOLTS/DIV*  
*400 KHZ*

*[Signature]*