

PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVK-2-17*

S.N.: *649*

Date: *OCT 14 81*

a) Output signal amplitude:

*≥ 0 TO +350 V*

b) Pulse width:

*0 TO 100 NS*

c) Rise time:

*≤ 1.0 NS*

d) Fall time:

*≤ 2.0 NS*

e) PRF:

*0 TO 5 KHZ*


f) Jitter, stability:

*OK*

g) Prime power:

*+15 ± 1.0 VOLT*

*550 mA MAX*



*10 NS/DIV*  
*± 350 VOLTS/DIV*  
*2 KHZ*

*[Signature]*