


PULSE GENERATOR
PERFORMANCE CHECK

Model: **AVK-A-17**
S.N.: **657**
Date: **NOV 23 81**



a) Output signal amplitude:

0 to +15V

b) Pulse width:

3 to 100 nsec,

c) Rise time:

≤ 100 psec

d) Fall time:

≤ 2 nsec

e) PRF:

500 F162

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 V.

225 mA

1.0 nsec/div
10 volts/div
400 kHz.