


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AUL-AU-P*

S.N.: *659*

Date: *NOV 20 81*



a) Output signal amplitude:

*0.70 ± 250 V.*

b) Pulse width:

*0 TO 100 NSEC*

c) Rise time:

*≤ 5 NSEC*

d) Fall time:

*≤ 5 NSEC*

e) PRF:

*0 TO 5 KHZ*

f) Jitter, stability:

*OK*

g) Prime power:

*+15 ± 1.0 VOLT*

*400 MA.*

*10 NSEC/DIV*

*320 VOLT/DIV*

*2 KHZ.*