


PULSE GENERATOR  
PERFORMANCE CHECK

Model: **AVI-MP-P**

S.N.: **668**

Date: **JAN 26 82**



- a) Output signal amplitude: **7, +40V**
- b) Pulse width: **0.5 TO 100 NSEC**
- c) Rise time: **500 P16 1+2**  
**≤ 0.5 NSEC**
- d) Fall time: **≤ 2.0 NSEC**
- e) PRF: **500 P16 2.**
- f) Jitter, stability: **OK**
- g) Prime power:

**1.0 NSEC/DIV.**

**32 VOLTS/DIV.**

**400 KHZ.**

**+15 ± 1.0 VOLT**

**185 mA**

**MAX**