


PULSE GENERATOR

PERFORMANCE CHECK

Model: **AVO-3A-P**


S.N.: **669**

Date: **JAN 26 82**



- a) Output signal amplitude: **25 to 75 Amps.**
- b) Pulse width: **2.5 NS (HALF POWER)**
- c) Rise time: **NA**
- d) Fall time: **NA**
- e) PRF: **0 TO 20 KHz**
- f) Jitter, stability: **OK**
- g) Prime power: **+15 ± 1.0 V**
250 mA.

$R_L = 5.1 \Omega$ 20 AMPS/DIV
5 NS/DIV



$R_L = 0.5 \Omega$ 200 AMPS/DIV
5 NS/DIV
20 KHz

[Handwritten signature]