


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AV11-5-1-P*

S.N.: *674*

Date: *NOV 13 81*




*10 VOLTS/DIV
1.0 NSEC/DIV
400 KHZ*

- a) Output signal amplitude:
≥ 10 VOLTS
- b) Pulse width:
≤ 135 PSEC
- c) Jitter, stability:
OK.
- d) Prime power:

+15 ± 1.0 V.

240 mA



*10 VOLTS/DIV.
100 PSEC/DIV.
400 KHZ*

A handwritten signature in dark ink, appearing to be 'JHC'.