


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-1-N-05-TTC*

S.N.: *678*

Date: *NOV 6 81*



10 VOLTS/DIV
1.0 NS/C/DIV
400 KHZ

Output signal amplitude:
0 TO -10V
Pulse width:
0 TO 4 NS/C
Rise time:
≤ 100 NS/C
Fall time:
≤ 135 NS/C
PRF:
0 TO 1.0 MHz
Jitter, stability:
OK.
Prime power:

+15 ± 1.0 V
310 mA (MAX)