


PULSE GENERATOR

PERFORMANCE CHECK

Model: **AVM-2-N**

S.N.: **684**

Date: **JAN 29 82**



a) Output signal amplitude:

0 TO 15V

b) Pulse width:

0 TO 2 μ SEC

c) Rise time:

≤ 100 μ SEC

d) Fall time:

≤ 135 μ SEC

e) PRF:

0 TO 25 μ Hz

f) Jitter, stability:

OK

g) Prime power:

+24 \pm 1.0 V

550 mA (MAX)

10V/DIV

100 μ SEC/DIV

4.0 μ Hz

PW_{0.5} = 25 μ SEC