


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVN-5-1-J*

S.N.: *687*

Date: *DEC 7 81*



1.0 NS/Div

5.0 VOLTS/Div

200 ns/Div

- a) Output signal amplitude:
0 TO +5V
- b) Pulse width:
0.3 TO 1.0 NS
- c) Rise time:
≤ 100 psec
- d) Fall time:
≤ 100 psec
- e) PRF:
50 - 260 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+24V, 450mA.

b) INPUT SIGNAL:
0 TO 1 V RMS.