


PULSE GENERATOR

PERFORMANCE CHECK

Model: **AUL-2-PS-C**

S.N.: **688**

Date: **DEC 15 81**



10 NSEC/DIV
320 VOLTS/DIV
1 KHZ.

a) Output signal amplitude:

0 TO +350 V

b) Pulse width:

0 TO 100 NSEC.

c) Rise time:

≤ 1 NSEC

d) Fall time:

≤ 2 NSEC

e) PRF:

0 TO 5 KHZ

f) Jitter, stability:

OK

g) Prime power:


120/240 V

50-60 HZ

PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVL-2-P-PS-C*
S.N.: *688 (MODIFIED & REPAIRED
OCT 1982)*
Date: *OCT 4 1982*



*320 VOLTS/DIV
10 NS/C DIV
2 KHz*

- a) Output signal amplitude:
0 TO +350 V
- b) Pulse width:
0 TO 100 NSCC.
- c) Rise time:
1 NSCC
- d) Fall time:
2 NSCC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
*110/220 V
50-60 Hz*