


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-P-EN-D*

S.N.: *695*

Date: *DEC 10 81*



- a) Output signal amplitude:
0 TO +5V
- b) Pulse width:
0 TO 6 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV.
1 NSEC/DIV.
1.0 MHz.

+24 ± 1.0 VOLT
370 mA
(MAX)

