


PULSE GENERATOR

PERFORMANCE CHECK

Model: **AUM-1-P-EW-12**

S.N.: **698**

Date: **DEC 10 81**



a) Output signal amplitude:

0 TO +5 V

b) Pulse width:

0 TO 6 NS

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 135 PSEC

e) PRF:

0 TO 25 MHz

f) Jitter, stability:

OK

g) Prime power:

+24 ± 1.0 VOLT

365 mA MAX.

5 VOLTS/DIV.

1.0 NS/DIV.

1.0 MHz.

