


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-1-P-EW-D*

S.N.: *700*

Date: *DEC 10 1981*



a) Output signal amplitude:

*0 TO +5 V*

b) Pulse width:

*0 TO 6 NS*

c) Rise time:

*≤ 100 PS*

d) Fall time:

*≤ 135 PS*

e) PRF:

*0 TO 25 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*+24 ± 1.0 VOLTS*

*340 mA (MAX)*

