

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-P-EW-D*

S.N.: *701*

Date: *DEC 10 81*

a) Output signal amplitude:

0 TO +5 VOLTS

b) Pulse width:

0 TO 6 NSEC

c) Rise time:

≤ 100 NSEC

d) Fall time:

≤ 135 PSEC

e) PRF:

0 TO 25 MHz

f) Jitter, stability:

OK

g) Prime power:

+24 ± 1.0 VOLTS,

350 mA (MAX)



