


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: **AUH-S-N**

S.N.: **702**

Date: **DEC 23 81**



a) Output signal amplitude:
 $\leq 15 \text{ V}$ ($R - 30 \text{ V}$)

b) Pulse width:
 $\leq 500 \text{ nsec}$

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0 \text{ V}$

130 mA

16 VOLTS / DIV

1.0 nsec / DIV

400 kHz