


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AV14-3-N*

S.N.: *703*

Date: *DEC 23 81.*



a) Output signal amplitude:
±15V (≈ -30V)

b) Pulse width:
≤ 500 psec

c) Jitter, stability:
OK

d) Prime power:

+15 ± 1 VOLT

180 mA

16 VOLTS/DIV.

1.0 nsec/DIV.

400 kHz.