


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: **AVG-2**

S.N.: **704**

Date: **NOV 20 81**



a) Output signal amplitude:

$\geq 140 \text{ V}$

b) Pulse width:

$\leq 2 \text{ nsec}$

c) Jitter, stability:

OK

d) Prime power:

+15 V

350 mA

(MAX)

100 V/DIV

2 nsec/DIV

20 KHZ