


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-5-1-A-P*

S.N.: *719*

Date: *JAN 14 82*



a) Output signal amplitude:

0 TO +10V

b) Pulse width:

≤ 130-150 PSEC.

c) Jitter, stability:

OK

d) Prime power:

+15 ± 1.0 V

200mA MAX.

*1.0 NSEC/DIV
10 VOLTS/DIV
400 KHZ.*