


PULSE GENERATOR
PERFORMANCE CHECK

Model: **AVO-2-P**

S.N.: **721**

Date: **JAN 14 1982**



- a) Output signal amplitude:
0 to +10 AMPS
- b) Pulse width:
0 TO 15 USEC.
- c) Rise time:
≤ 1.0 USEC
- d) Fall time:
≤ 1.0 USEC.
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

10 AMP/DIV
5 USEC/DIV
20 KHZ

+15 ± 1.0 V
250 mA (MAX)

[Handwritten signature]