


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AUM-1-P-D*

S.N.: *723*

Date: *JAN 14 1982*



- a) Output signal amplitude:  
*0 TO +5 VOLTS*
- b) Pulse width:  
*0 TO 6 NS*
- c) Rise time:  
*≤ 100 PSEC.*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+24 ± 1.0 VOLT*

*350 mA (MAX)*

*5 VOLTS/DIV*

*1.0 NS/DIV*

*4 MHz.*