


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-3-P-D-PS*

S.N.: *726*

Date: *JAN 29 82*



- a) Output signal amplitude:
0 TO +15V
- b) Pulse width:
0 TO 15 NS
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 600 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

10 V/DIV
1 NS/DIV
4 MHz
PW_{IN} = 25 NS

110/220V
50-60 Hz