


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: **AVH -MPT.\***

S.N.: **727**

Date: **FEB 9 82**



- a) Output signal amplitude:  
**≤ -30 VOLTS**
- b) Pulse width:  
**≥ 2 NSCC (FWHM)**
- c) Jitter, stability:  
**OK**
- d) Prime power:

**+15 ± 1.0 VOLT.**

**150 mA (MAX)**

**2 NSCC / DIV.**

**32 VOLTS / DIV.**

**400 KHz**

**\* MODIFIED  
FOR PW =  
2 NSCC (FWHM)**