


PULSE GENERATOR
PERFORMANCE CHECK


Model: **AVK-V-N-TTL-05**

S.N.: **733**

Date: **JAN 19 82**



1 NSEC/DIV
10 VOLTS/DIV
40 KHZ



10 NSEC/DIV
10 VOLTS/DIV
40 KHZ

a) Output signal amplitude:

≥ -5 V

b) Pulse width:

0 TO 50 NSEC

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 2 NSEC

e) PRF:

0 TO 50 KHZ

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1 VOLT

260 mA (MAX)