


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: **AVH-5-1-A 7**

S.N.: **734**

Date: **FEB 9 82**



a) Output signal amplitude:

0 TO +10V

b) Pulse width:

≤ 135 nsec.

c) Jitter, stability:

OK

d) Prime power:

+15 ± 1 Volt

150 mA (max)

10 Volts/DIV

1.0 nsec/DIV.

400 kHz.