


PULSE GENERATOR

PERFORMANCE CHECK

Model: **AUP-AU-1-P**

S.N.: **742**

Date: **MAR 1 82**



- a) Output signal amplitude:
0 TO +10V
- b) Pulse width:
0 TO 4 NSOR
- c) Rise time:
≤ 100 PSOR
- d) Fall time:
≤ 135 PSOR
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLT/DIV

1 NSOR/DIV

400 KHz

+15 ± 1.0V

270 mA