


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV0-2-P*

S.N.: *744*

Date: *JAN 14 1982*



- a) Output signal amplitude:
0 TO +10 AMPS
- b) Pulse width:
0 TO 15 NSEC
- c) Rise time:
≤ 1.0 NSEC
- d) Fall time:
≤ 1.0 NSEC
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

10 AMP/DIV
5 NSEC/DIV
20 KHZ.

+15 ± 1.0 V
220 mA (MAX)

[Signature]